

MAKLUMAT SPESIFIKASI ITEM

www.fineprocurementplus.uitm.edu.my



MAKLUMAT SPESIFIKASI ITEM

No Pelawaan	: UiTM/B6/PER/T/B/0325/0006
Butiran Tawaran	: CADANGAN MEMBEKAL DAN MENGHANTAR BARANGAN HAEMODIALISIS UNTUK KEGUNAAN JABATAN
Keterangan Item	: 0610020100052 : Catheter / Guidewire
Nama Spesifikasi	: Bloodline for Haemodialysis
Kuantiti	: 6500.000

Butiran Keperluan	Calit	Tawaran Syarikat
1 TITLE: Bloodline for heamodialysis		
I Use: Bloodline for heamodialysis and haemodiafiltration treatment	☐	
2 Discription		
I Compatibility with machine interfaces is an indispensable prerequisite for a safe dialysis treatment. Ensure interface tests on the blood tubing systems and dialysis machines forthe highest safety and performance	☐	
II ETO-sterilized lines that undergo a special pre-conditioning and degassing process to ensure a safe bloodline with high biocompatibility.	☐	
III Beta-sterilization, also known as e-beam sterilization, offers excellent biocompatibility. E-beam sterilization shows fewer material alterations in comparison to gamma-sterilization	☐	
IV Saline free priming of extracorporeal circuit	☐	

Butiran Keperluan

- V Exact measurement of arterial pressure via pressure dome.
- VI Trouble-free air detection directly on system tube.
- VII Clear layout and optimised tube lengths
- VIII Machine assisted set-up and disassembling
- IX Reduced filling volume and less blood-air contact
- 3 Expiry date should not less than 1 years from delivery date.
- 4 Compatible with 5008s MACHINE
- 5 packaging: 1unit/pack

Calit	Tawaran Syarikat
☐	
☐	
☐	
☐	
☐	