

MAKLUMAT SPESIFIKASI ITEM

www.fineprocurementplus.uitm.edu.my



MAKLUMAT SPESIFIKASI ITEM

No Pelawaan	: UiTM/B6/PER/T/B/0325/0006
Butiran Tawaran	: CADANGAN MEMBEKAL DAN MENGHANTAR BARANGAN HAEMODIALISIS UNTUK KEGUNAAN JABATAN
Keterangan Item	: 0609040200003 : Lab Supplies - Membrane Filters
Nama Spesifikasi	: HI-FLUX DIALYZER FOR HEMODIALYSIS
Kuantiti	: 7500.000

Butiran Keperluan

1 HI-FLUX DIALYZER :

- I Clearance ($Q_B = 300$ mL/min)
- II Cytochrome c = At least 89 or more
- III Inulin = At least 113 or more
- IV Vitamin B₁₂ = At least 185 or more
- V Phosphate = At least 244 or more
- VI Creatinine = At least 259 or more
- VII Urea = At least 279 or more

Calit

Tawaran Syarikat

☐☐☐☐☐☐☐

Butiran Keperluan

VIII Sieving coefficients :

- A) Albumin : At least 0.001 or more
- B) Myoglobin = At least 0.1 or more
- C) $\beta 2$ - microglobulin = At least 0.7 or more
- D) Inulin = At least 1 or more
- E) Effective surface (m_2) = At least 1.8 or more
- F) K_0A Urea = At least 1,394 or more
- G) Priming volume (mL) = At least 95 or more

IX Manufacturing to defect

- A) Defect items can be returnable and refundable

2 expiry date atleast 1 year from date delivery

3 Packaging: 1unit in 1 pack

Calit	Tawaran Syarikat
☐	
☐	
☐	
☐	
☐	
☐	
☐	
☐	
☐	
☐	