

MAKLUMAT SPESIFIKASI ITEM

www.fineprocurementplus.uitm.edu.my



MAKLUMAT SPESIFIKASI ITEM

No Pelawaan	: UiTM/B6/PER/T/B/0325/0006
Butiran Tawaran	: CADANGAN MEMBEKAL DAN MENGHANTAR BARANGAN HAEMODIALISIS UNTUK KEGUNAAN JABATAN
Keterangan Item	: 0609040200003 : Lab Supplies - Membrane Filters
Nama Spesifikasi	: LOW FLUX DIALYSER
Kuantiti	: 600.000

Butiran Keperluan	Calit	Tawaran Syarikat
1 Title:LOW FLUX DIALYSER		
I Use:filter for hemodialysis treatment	☐	
2 Discription		
I Low flux dialyzers are an option for new patients or acute and chronic dialysis where a lower rate of fluid removal (e.g., ultrafiltration coefficient) is desired and desired Kt/V adequacy goals with the same reliability in performance and biocompatibility as other high flux dialyzers. low flux dialyzers provides clinicians with options to individualize patient care.	☐	
3 Performance data		
I Clearance: QB: (200ml/min)	☐	
A) Urea mmol/L193 Creatinine umol/L181	☐	

Butiran Keperluan

Calit

Tawaran Syarikat

- Phosphate mmol/L170
- Vitamin B12 pg/mL121
- Effective surface area (m²)1.8
- Blood filling volume (ml)95
- ApplicationHD
- 4 Expiry date should not less than 1 years from delivery date.
- 5 compatible with 5008s MACHINE
- 6 packaging: 1unit/pack