

MAKLUMAT SPESIFIKASI ITEM

www.fineprocurementplus.uitm.edu.my



MAKLUMAT SPESIFIKASI ITEM

No Pelawaan	: UiTM/B1/PER/SHR/P/0625/0012
Butiran Tawaran	: CADANGAN PEMBAIKAN 1 UNIT GAMMA SPECTROSCOPY SYSTEM DI MAKMAL ATOM DAN SINARAN BLOK
Keterangan Item	: 1101030900001 : Perkhidmatan Penyelenggaraan / Baikpulih / Servis - Mesin, Peralatan dan
Nama Spesifikasi	: PEMBAIKAN 1 UNIT GAMMA SPECTROSCOPY SYSTEM DI MAKMAL ATOM DAN SINARAN BLOK D
Kuantiti	: 1.000

Butiran Keperluan

1 DETECTOR MODEL GEM30P4-70-RB; S/N: 54-TP23665A

I Cryostat configuration: CFG-PV4

II Preamplifier Model 257P; S/N: 17067581

III H.V. Filter Model 138EMI; S/N: 17045682

2 Evaluation fee inclusive of freight charges to US and back to KL including with installation.

3 Complex Repair 40% GEM = 80% inclusive of crystal refurbish and vacuum refresh (Pump and Bake) and Preamp Repair.

4 Costing send along Mobius to repair (manufacturer in US)

I Evaluation fee inclusive of freight charges to US and back to KL including with installation.

Calit

☐☐☐☐

Tawaran Syarikat

--

Butiran Keperluan

- II Performing troubleshooting and full reliability test to electronics and mechanical parts (replacement of mother board).
- 5 Training and commisioning for complete training how to functioning the ALL system and software
- 6 Provide the standard operating manual for this equipment in English.
- 7 Site visit must be completed.
- 8 The system must be functioning properly and ready to use without any additional cost.

Calit

☐

Tawaran Syarikat