

MAKLUMAT SPESIFIKASI ITEM

www.fineprocurementplus.uitm.edu.my



MAKLUMAT SPESIFIKASI ITEM

No Pelawaan	: UiTM/M3/PER/SHR/P/0625/0002
Butiran Tawaran	: CADANGAN MENYELENGGARA PERALATAN MAKMAL UITM KAMPUS JASIN MELAKA 2025 FASA 2
Keterangan Item	: 0609030200055 : Rotavapor/Rotary Evaporator
Nama Spesifikasi	: FPA - MEMBAIKI DAN MENYELENGGARA ROTARY EVAPORATOR (PROTECH, 2016000008, MP2/B2)
Kuantiti	: 1.000

Butiran Keperluan

Calit

Tawaran Syarikat

- 1 To service all the glass parts (clean interior&exterior)
- 2 To ensure water is flowing for condensation process
- 3 To repair pump (replace if necessary)
- 4 To check all wiring (heater and pump)
- 5 To check any leakage